

**QwikCheck™ Beads Batch Release Form**

BATCH #	040423	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	April-24	LOT#	040423001	LOT#	040423002	LOT#	040423003	Conc	MSC
For use with:	Conc.			Conc.			Conc	MSC	
SQA / QwikCheck gold Automated	47	+/-	6.6	25	+/-	5.0	0.0	0.0	
SQA-V Visualization	50	+/-	12.5	25	+/-	6.3	0.0	NA	
Hemocytometer	51	+/-	12.8	24	+/-	6.0	0.0	NA	
MAKLER	50	+/-	12.5	26	+/-	6.5	0.0	NA	
Fixed Coverslip	51	+/-	12.8	25	+/-	6.3	0.0	NA	

Clinical research assistant

SZ

4-Apr-23

Signature

Date

Clinical Manager

F.C

4-Apr-23

Signature

Date

CEO

MØ

4-Apr-23

Signature

Date