

MEDICAL ELECTRONIC SYSTEMS

WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form											
BATCH #		011125		Level 1			Level 2			Negative Control	
EXP DATE:		November-26		LOT# 011125001			LOT# 011125002			LOT#	011125003
For use with:				Conc.			Conc.			Conc	MSC
SQA / QwikCheck gold Automated				48	+/-	6.7	26	+/-	5.2	0.0	0.0
Manual	SQA-V Visualization			48	+/-	12.0	24	+/-	6.0	0.0	NA
	Hemocytometer			49	+/-	12.3	25	+/-	6.3	0.0	NA
	MAKLER			49	+/-	12.3	25	+/-	6.3	0.0	NA
	Fixed Coverslip			48	+/-	12.0	24	+/-	6.0	0.0	NA