

QwikCheck™ Beads Batch Release Form									
BATCH #	110626	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	June-27	LOT#	110626001		LOT#	110626002		LOT#	110626003
For use with:		Conc.			Conc.			Conc	MSC
SQA / QwikCheck Gold Automated		47	+/-	6.6	26	+/-	5.2	0.0	0.0
SQA-V Visualization		48	+/-	12.0	26	+/-	6.5	0.0	NA
Hemocytometer		48	+/-	12.0	26	+/-	6.5	0.0	NA
MAKLER		48	+/-	12.0	26	+/-	6.5	0.0	NA
Fixed Coverslip		47	+/-	11.8	26	+/-	6.5	0.0	NA

S.L.
Laboratory Technician

M.Y.
Clinical Director

S.L.
Signature

M.Y.
Signature

25-Jun-26
Date

25-Jun-26
Date