

Work Instruction No: WI1-030



QwikCheck[™] Liquefaction Kit Batch Release Form

QwikCheck [™] Liquefaction Kit produced:	October	2022
	Month	Year

Chymotrypsin activity = 1200 units/mg enzyme

Notes:

- PASS criteria: Mean ≥1200 units/mg enzyme
- Certificate of Analysis (Faizyme) for each Lot Number should be attached to this form.

Number of kits produced	187
Chymotrypsin QC Approval date	15 th September 2022
Product Catalog number	0900; 20 vials

Batch number: 220809_1022

Lot Number (Faizyme): 220809

Exp. Date: OCT 2023 Month Year

> <u>Victoria Goldenberg</u> Clinical Research Assistant

02nd November 2022 Date