

QwikCheck™ Beads Batch Release Form									
BATCH #	090426	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	April-27	LOT#	090426001	LOT#	090426002	LOT#	090426003		
For use with:		Conc.			Conc.			Conc	MSC
SQA / QwikCheck Gold Automated	49	+/-	6.9	27	+/-	5.4	0.0	0.0	
SQA-V Visualization	48	+/-	12.0	27	+/-	6.8	0.0	NA	
Hemocytometer	48	+/-	12.0	26	+/-	6.5	0.0	NA	
MAKLER	48	+/-	12.0	26	+/-	6.5	0.0	NA	
Fixed Coverslip	48	+/-	12.0	27	+/-	6.8	0.0	NA	

\_\_\_\_\_  
S.L.  
Laboratory Technician

\_\_\_\_\_  
*S.L.*  
Signature

\_\_\_\_\_  
23-Apr-26  
Date

\_\_\_\_\_  
M.Y.  
Clinical Director

\_\_\_\_\_  
*M.Y.*  
Signature

\_\_\_\_\_  
23-Apr-26  
Date