

MEDICAL ELECTRONIC SYSTEMS  
WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #	011124	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	November-25	LOT#	011124001		LOT#	011124002		LOT#	011124003
For use with:		Conc.			Conc.			Conc	MSC
SQA / QwikCheck gold Automated		48	+/-	6.7	25	+/-	5.0	0.0	0.0
SQA-V Visualization		48	+/-	12.0	25	+/-	6.3	0.0	NA
Hemocytometer		48	+/-	12.0	26	+/-	6.5	0.0	NA
MAKLER		48	+/-	12.0	25	+/-	6.3	0.0	NA
Fixed Coverslip		47	+/-	11.8	25	+/-	6.3	0.0	NA

\_\_\_\_\_  
D.K  
Laboratory Technician

\_\_\_\_\_  
S.Z  
Production Laboratory Manager

\_\_\_\_\_  
*Dk*  
Signature

\_\_\_\_\_  
*SZ*  
Signature

\_\_\_\_\_  
11-Nov-24  
Date

\_\_\_\_\_  
11-Nov-24  
Date