

MEDICAL ELECTRONIC SYSTEMS

WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #	010725	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	July-26	LOT#	010725001		LOT#	010725002		LOT#	010725003
For use with:		Conc.			Conc.			Conc	MSC
SQA / QwikCheck Gold Automated		47	+/-	6.6	25	+/-	5.0	0.0	0.0
SQA-V Visualization		49	+/-	12.3	26	+/-	6.5	0.0	NA
Hemocytometer		49	+/-	12.3	25	+/-	6.3	0.0	NA
MAKLER		49	+/-	12.3	25	+/-	6.3	0.0	NA
Fixed Coverslip		48	+/-	12.0	25	+/-	6.3	0.0	NA

D.K

Laboratory Technician

D.K

Signature

8-Jul-25

Date

S.Z

Production Lab Manager

S.Z

Signature

8-Jul-25

Date