Appendix 2
COA and Batch Release Form

Batch Details			
Product Name	**QwikCheck™ DFI Kit**	**Batch Number**	010626
Product Number	A-CA-01873-00 / EU-A-CA-01940-00	**Expiration date**	JUN 2027
Storage Conditions	2°C-8°C (36°F-46°F)	**Number of Kits**	35

Visual and pH QC of Bulk materials			
Kit Component	**Test**	**Pass Criteria (test specifications)**	**Results Pass/Fail**
Lys 1 Solution	Turbidity	Clear	Pass
	Color	Colorless	Pass
	pH	7.5 +/- 0.5	Result: 7.58
			Pass
Lys 2 Solution	Turbidity	Clear	Pass
	Color	Colorless	Pass
	pH	10.0 +/- 0.5	Result: 10.00
			Pass
Lys 2 Powder	Color	White	Pass
Diluent Solution Washing Solution	Turbidity	Clear	Pass
	Color	Colorless	Pass
	pH	7.0-7.5	Result: 7.25
			Pass
Fixative	Turbidity	Clear	Pass
	Color	Colorless	Pass
Giemsa A	Homogeneity	Homogenous, no precipitation	Pass
	Color	Dark Blue	Pass
Giemsa B	Turbidity	Clear	Pass
	Color	Colorless	Pass
Performance Testing			
Test	**Definition**		**Results Pass/Fail**
Functional Performance	Ability to complete a test run using the DFI kit components of the specified batch (Appendix 3)		**Pass**

This document confirms that these parameters have been tested in a QC laboratory and have been satisfactorily completed.

S.L	03-June-2026
Laboratory Technician	**Date**
M.Y	03-June-2026
Clinical Director	**Date**