

MEDICAL ELECTRONIC SYSTEMS
WWW.MES-GLOBAL.COM

QwikCheck™ Low Range Beads Batch Release Form									
BATCH #	220625	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	Mar-26	LOT#	220625001		LOT#	220625002		LOT#	220625003
For use with:		Conc.			Conc.			Conc.	MSC
SQA / QwikCheck gold Automated		15	+/-	6.0	4	+/-	2.6	0.0	0.0
SQA-V Visualization		16	+/-	6.4	4	+/-	2.6	0.0	NA
Fixed Coverslip		16	+/-	6.4	3	+/-	2.0	0.0	NA
Hemocytometer		15	+/-	6.0	NA	+/-	NA	0.0	NA

D.K, Production Laboratory Team Lead

D.K

Signature

25-Jun-25

Date

S.Z, Production Laboratory Manager

SZ

Signature

25-Jun-25

Date