



# Comparison Of Manual Vs Automated (SQA Vision) Semen Analysis: A Double Blind Prospective Study

Praveen Kumar, Omwati Vats, Sarman Singh, Lev Rabinovitch\*

Division of Clinical Microbiology and Molecular Medicine, Department of Laboratory Medicine, All India Institute of Medical Sciences, New Delhi, India

- ❖ Manual semen analysis suffers from analytical variability & subjective variation.
- ❖ Manual assessment of sperm morphology is associated with difficulties related to lack of objectivity, variation in interpretation or poor performance in external quality-control assessments.

- ❖ To compare an automated computerized semen analysis system (SQA-Vision) to the conventional manual method in terms of accuracy and precision.

- ❖ Type of study: Prospective (Double blind), analytical
- ❖ Conducted in Division of Clinical Microbiology and Molecular Medicine, Department of Lab Medicine, AIIMS, New Delhi between July- September 2016.
- ❖ 250 fresh semen samples were tested by both, manual method & SQA-Vision, in duplicate at room temperature.
- ❖ Manual testing was done by 2 independent operators & 1 manual operator ran SQA-Vision immediately following the manual analysis of motility to prevent bias.
- ❖ Sperm concentration, total and progressive motility & morphology were assessed according to the 5<sup>th</sup> WHO semen testing laboratory manual, 2010.
- ❖ Statistical analysis of data was done using MedCalc (Belgium) and Excel programs.


- ❖ Comparison of the SQA-Vision & manual results is presented in the table 1 below.

**Table 1: Comparison of SQA-Vision & manual results of semen analysis**


Statistical Parameters		Semen Parameters			
		Sperm Conc.	Total Motile PR + NP	Progressive Motile PR	Morph. Normal Forms
SQA-Vision	Mean ( $\times 10^6/ml$ )	45.4	30.0	20.8	6.4
	CV, %	3.1	5.6	4.4	5.3
Manual	Mean, $\times 10^6/ml$	46.3	35.0	24.5	7.7
	CV, %	11.0	14.0	19.6	17.8
	Sensitivity (%)	100.0	95.2	95.9	100.0
	Specificity (%)	99.3	89.6	85.0	98.9
	Concordance	0.97	0.84	0.86	0.84
	Correlation	0.97	0.87	0.89	0.89
	Pearson Correlation	0.97	0.87	0.89	0.89
	Bias Correction (accuracy)	1.0	0.96	0.97	0.95

- ❖ CVs for SQA-Vision are much lower than for manual analysis which demonstrates that SQA-Vision's precision is higher.


**Fig 1. Sperm Concentration**




**Fig 2. Sperm Total motility**



**Fig 3. Progressive motility**



**Fig 4. Sperm Normal Morphology**



- ❖ SQA-Vision & manual semen parameters' mean values are quite close demonstrating no systematic discrepancies.
- ❖ Sensitivity, specificity, concordance & correlation coefficients are very high indicating a high level of accuracy & close agreement between the 2 methods.
- ❖ Bias correction coefficients are between 0.95 & 1.0 for different semen parameters showing a high level of agreement.

- ❖ The automated semen analyzer SQA-Vision is faster and provides a higher level of standardization and precision vs. manual semen assessment.
- ❖ The simplicity of operating the automated SQA-Vision minimizes the need for highly skilled professionals.

- ❖ We acknowledge the technical support of Mr Alayamani Kannan and technical staff of Laboratory Medicine for their immense support in the present work.