

MEDICAL ELECTRONIC SYSTEMS

WWW.MES-GLOBAL.COM

QwikCheck™ Low Range Beads Batch Release Form									
BATCH #	091225	Level 1			Level 2			Negative Control	
EXP DATE:	Sep-26	LOT#	091225001	LOT#	091225002	LOT#	091225003		
For use with:		Conc.			Conc.			Conc	MSC
SQA / QwikCheck gold Automated		16	+/-	6.4	4	+/-	2.6	0.0	0.0
Manual	SQA-V Visualization	16	+/-	6.4	3	+/-	2.0	0.0	NA
	Fixed Coverslip	15	+/-	6.0	3	+/-	2.0	0.0	NA
	Hemocytometer	16	+/-	6.4	NA	+/-	NA	0.0	NA