

MEDICAL ELECTRONIC SYSTEMS
WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #	010126	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	January-27	LOT#	010126001	LOT#	010126002	LOT#	010126003		
For use with:		Conc.			Conc.			Conc	MSC
SQA / QwikCheck Gold Automated		48	+/-	6.7	25	+/-	5.0	0.0	0.0
SQA-V Visualization		49	+/-	12.3	25	+/-	6.3	0.0	NA
Hemocytometer		48	+/-	12.0	25	+/-	6.3	0.0	NA
MAKLER		49	+/-	12.3	25	+/-	6.3	0.0	NA
Fixed Coverslip		48	+/-	12.0	25	+/-	6.3	0.0	NA

S.L
Laboratory Technician

S.L
Signature

7-Jan-26
Date

D.K
Production Lab Assistant Manager

M
Signature

7-Jan-26
Date