

	Medical Electronic Systems Ltd	Effective Date: 20 Jan, 2025
Version Number: 7	W11-086 QwikCheck TM Vitality Kit Production File	Page 5 of 7

Appendix 1

QwikCheck™ Vitality Kit Batch Release Form

QwikCheck™ Vitality Kit produced: 11 MAY 2025
 Day Month Year

Batch number: #1057_110525

Number Of kits produced: 80

Exp. Date: MAY 2026
 Month Year

Components Lot # & Exp. Dates:

Material	LOT #	Exp. date
NaCl	240670	29/06/2027
Eosin Y	SHBR8226	N/A

QC results:

Sample #	% Motility (Natural slides)	% Vitality (Stained slides)
1- Fresh	62	92
2- Frozen	2.9	10

Pass / Fail PASS

S.Z
 Production Laboratory Manager

Date: 11th May 2025