QwikCheck™ Low Range Beads Batch Release Form										
BATCH #	**111223**	**Level 1**			**Level 2**			**NEGATIVE CONTROL**	
EXP DATE:	**Sep-24**	**LOT#**	**111223001**		**LOT#**	**111223002**		**LOT#**	**111223003**
For use with:		**Conc.**			**Conc.**			**Conc**	**MSC**
SQA / QwikCheck gold Automated		**14**	**+/-**	**5.6**	**3**	**+/-**	**2.0**	**0.0**	**0.0**
SQA-V Visualization		**15**	**+/-**	**6.0**	**4**	**+/-**	**2.6**	**0.0**	**NA**
Fixed Coverslip		**14**	**+/-**	**5.6**	**3**	**+/-**	**2.0**	**0.0**	**NA**
Hemocytometer		**14**	**+/-**	**5.6**	**NA**	**+/-**	**NA**	**0.0**	**NA**

K.K, Laboratory Technician	K.K	18-Dec-23
	Signature	Date
S.Z, Production Lab Lead Tech	S.Z	18-Dec-23
	Signature	Date
M.D, CEO	Marcia Deutsch	18-Dec-23
	Signature	Date