

Certificate of Analysis

For Invitro diagnostics use only. Not for human/therapeutic procedures.

Trade Name Product Name Product Number Product Lot	Earle's Balanced Salt Solution QwikCheck [™] Dilution; 50ml 06-2010-02-1E 242700712_231224	
Storage Temp	15-30°C	
Test	SPECIFICATION	RESULTS
рН	7.3-7.7	7.475
Osmolality	270 - 300 mOsm/kg	290 mOsm/kg
Endotoxins	Check and record	<0.100 EU/ml
Sterility testing	Negative	Pass test
Appearance	Clear solution	Pass test

Expiration Date:

DEC/2025

This document confirms that these parameters have been tested in a QC laboratory and have been satisfactorily completed.

Signature:

Production Laboratory Manager

Date: 23 December 2024

MES-0188-04