

QwikCheck™ Low Range Beads Batch Release Form

BATCH #	111223	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	Sep-24	LOT#	111223001	LOT#	111223002	LOT#	111223003		
For use with:		Conc.			Conc.			Conc	MSC
SQA / QwikCheck gold Automated		14	+/-	5.6	3	+/-	2.0	0.0	0.0
SQA-V Visualization		15	+/-	6.0	4	+/-	2.6	0.0	NA
Fixed Coverslip		14	+/-	5.6	3	+/-	2.0	0.0	NA
Hemocytometer		14	+/-	5.6	NA	+/-	NA	0.0	NA

K.K, Laboratory Technician

K.K

18-Dec-23

Signature

Date

S.Z, Production Lab Lead Tech

S.Z

18-Dec-23

Signature

Date

M.D, CEO

Marcia Deutsch

18-Dec-23

Signature

Date