



Medical Electronic Systems Ltd.
A-Tech (a division of M.E.S.)
20 Alon Hatavor Street, Caesarea
Industrial Park 38900, Israel
E-mail: mes@mes-ltd.com

Certificate of Analysis

For Invitro diagnostics use only. Not for human/therapeutic procedures.

Trade Name	Earle's Balanced Salt Solution
Product Name	QwikCheck™ Dilution; 50ml
Product Number	06-2010-02-1E
Product Lot	2209192_240322
Storage Temp	15-30°C

Test	SPECIFICATION	RESULTS
pH	7.3-7.7	7.515
Osmoiality	270 - 300 mOsm/kg	283 mOsm/kg
Endotoxins	Check and record	<0.005 EU/ml
Sterility testing	Negative	Pass test
Appearance	Clear solution	Pass test

Expiration Date: MAR/2023

This document confirms that these parameters have been tested in a QC laboratory and have been satisfactorily completed.

Signature: _____
Clinical Manager

Date: 24th MAR 2022