



Medical Electronic Systems Ltd.  
A-Tech (a division of M.E.S.)  
20 Alon Hatavor Street, Caesarea  
Industrial Park 38900, Israel  
E-mail: [mes@mes-ltd.com](mailto:mes@mes-ltd.com)

## Certificate of Analysis


For Invitro diagnostics use only. Not for human/therapeutic procedures.

Trade Name	<b>Earle's Balanced Salt Solution</b>
Product Name	<b>QwikCheck™ Dilution; 50ml</b>
Product Number	06-2010-02-1E
Product Lot	2126578_181021
Storage Temp	15-30°C

<b>Test</b>	<b>SPECIFICATION</b>	<b>RESULTS</b>
pH	7.3-7.7	7.48
Osmolality	270 - 300 mOsm/kg	282 mOsm/kg
Endotoxins	Check and record	<0.005 EU/ml
Sterility testing	Negative	Pass test
Appearance	Clear solution	Pass test

Expiration Date                      OCT 2022

This document confirms that these parameters have been tested in a QC laboratory and have been satisfactorily completed.

Signature:   
Clinical manager

Date: \_\_\_\_ OCT 2021 \_\_\_\_