



Title: QwikCheck™ Liquefaction Kit Production File

Work Instruction No: WI1-030

Revision: 19

Page 1 of 1



QwikCheck™ Liquefaction Kit Batch Release Form

QwikCheck™ Liquefaction Kit produced: MAY 2022
Month Year

Chymotrypsin activity = 1200 units/mg enzyme

Notes:

- PASS criteria: Mean ≥ 1200 units/mg enzyme
- Certificate of Analysis (Faizyme) for each Lot Number should be attached to this form.

Number of kits produced	404
Chymotrypsin QC Approval date	07 th March 2022
Product Catalog number	0900; 20 vials

Batch number: 211216 0522

Lot Number (Faizyme): 211216

Exp. Date: MAY 2023
Month Year

Frida Cohen
Clinical Research Assistant

29th May 2022
Date