



Title: QwikCheck™ Liquefaction Kit Production File

Work Instruction No: WI1-030

Revision: 19

Page 1 of 1



## QwikCheck™ Liquefaction Kit Batch Release Form

QwikCheck™ Liquefaction Kit produced: June 2022  
Month Year

Chymotrypsin activity = 1200 units/mg enzyme

**Notes:**

- PASS criteria: Mean  $\geq 1200$  units/mg enzyme
- Certificate of Analysis (Faizyme) for each Lot Number should be attached to this form.

Number of kits produced	<b>241</b>
Chymotrypsin QC Approval date	26 <sup>th</sup> May 22
Product Catalog number	0900; 20 vials

Batch number: 220303 0622

Lot Number (Faizyme): 220303

Exp. Date: JUN 2023  
Month Year

Frida Cohen  
Clinical Research Assistant

06<sup>th</sup> June 2022  
Date