

QwikCheck™ Low Range Beads Batch Release Form

BATCH #	241121	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	30-Aug-22	LOT#	241121001	LOT#	241121002	LOT#	241121003		
For use with:		Conc.			Conc.			Conc	MSC
SQA / QwikCheck™ gold Automated		14	+/-	5.6	3	+/-	2.0	0.0	0.0
SQA-V Visualization		14	+/-	5.6	3	+/-	2.0	0.0	NA
Fixed Coverslip		15	+/-	6.0	4	+/-	2.6	0.0	NA
Hemocytometer		14	+/-	5.6	NA	+/-	NA	0.0	NA

Noa Zecharia

Clinical research assistant



Signature

28-Nov-21

Date

Frida Cohen, Clinical Manager



Signature

28-Nov-21

Date

Marcia Deutsch, CEO



Signature

28-Nov-21

Date