

Technical Release Bulletin
QwikCheck™ Beads Open Vial Shelf Life
Issue date: April 11, 2018

Background:

QwikCheck™ Beads (Medical Electronic Systems) have an open and a closed vial shelf life. This labeling insures accurate results that could be compromised due to bead evaporation over time which can be related to high ambient laboratory temperature and/or infrequent use of the beads.

Storage conditions:

The stability of the QwikCheck™ Beads control material is highly dependent on maintaining an intact matrix volume. Therefore, special storage conditions and timeframes for use after opening are recommended as follows:

1. Each kit contains two known concentrations of **QwikCheck™ Beads** in two 5 ml aliquots and one 5 ml negative concentration and motility control.
2. Store unopened bead kits @ 15-30°C (60-86°F).
3. After opening for the first time, refrigerate @ 2-8°C (36-46°F) for 90 days or store @ 15-30°C (60-86°F) for 30 days.
4. Refrigerated bottles must be brought to room temperature before use.
5. The labeled storage expiration date assumes that **QwikCheck™ Beads** are stored in their original containers and are tightly capped to prevent evaporation.
6. QwikCheck™ beads are stable and show no loss of expected performance characteristics after transport/storage over a period of 72 hours at the temperature range of -20°C to +37°C (-4°F to +99°F).

Compliance Date: Effective April 11, 2018. **Authority:** Marina Friling, Clinical Manager
Distribution: ALL QwikCheck Bead users



www.mes-global.com

QwikCheck™ Beads Open Vial Shelf Life_11_April_18