


**QwikCheck™ Beads Batch Release Form**

BATCH #	140322	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	13-APR-23	LOT#	140322001	LOT#	140322002	LOT#	140322003		
For use with:		Conc.			Conc.			Conc.	MSC
SQA / QwikCheck gold Automated		47	+/-	6.6	25	+/-	5.0	0.0	0.0
SQA-V Visualization		48	+/-	12.0	25	+/-	6.3	0.0	NA
Hemocytometer		49	+/-	12.3	24	+/-	6.0	0.0	NA
MAKLER		50	+/-	12.5	25	+/-	6.3	0.0	NA
Fixed Coverslip		49	+/-	12.3	26	+/-	6.5	0.0	NA

Sapir Zrihan, Clinical research assistant



Signature

22-Mar-22

Date

Frida Cohen, Clinical Manager



Signature

22-Mar-22

Date

Marcia Deutsch, CEO

Mdeutsch

Signature

22-Mar-22

Date