

QwikCheck™ Beads Batch Release Form

BATCH #	060823	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	August-24	LOT#	060823001	LOT#	060823002	LOT#	060823003	Conc	MSC
For use with:	Conc.			Conc.			Conc	MSC	
SQA / QwikCheck gold Automated	48	+/-	6.7	25	+/-	5.0	0.0	0.0	
SQA-V Visualization	48	+/-	12.0	25	+/-	6.3	0.0	NA	
Hemocytometer	50	+/-	12.5	23	+/-	5.8	0.0	NA	
MAKLER	46	+/-	11.5	24	+/-	6.0	0.0	NA	
Fixed Coverslip	52	+/-	13.0	27	+/-	6.8	0.0	NA	

Clinical research assistant

S.B

16-Aug-23

Signature

Date

Clinical Manager

F.O

16-Aug-23

Signature

Date

CEO

Marcia Deutsch

16-Aug-23

Signature

Date