


### QwikCheck™ Beads Batch Release Form

| BATCH #                         | 100722  | Level 1 |           |      | Level 2   |      |           | NEGATIVE CONTROL |     |
|---------------------------------|---------|---------|-----------|------|-----------|------|-----------|------------------|-----|
| EXP DATE:                       | July-23 | LOT#    | 100722001 | LOT# | 100722002 | LOT# | 100722003 |                  |     |
| For use with:                   |         | Conc.   |           |      | Conc.     |      |           | Conc             | MSC |
| SQA / QwikCheck™ gold Automated |         | 48      | +/-       | 6.7  | 25        | +/-  | 5.0       | 0.0              | 0.0 |
| SQA-V Visualization             |         | 46      | +/-       | 11.5 | 24        | +/-  | 6.0       | 0.0              | NA  |
| Hemocytometer                   |         | 48      | +/-       | 12.0 | 24        | +/-  | 6.0       | 0.0              | NA  |
| MAKLER                          |         | 47      | +/-       | 11.8 | 25        | +/-  | 6.3       | 0.0              | NA  |
| Fixed Coverslip                 |         | 48      | +/-       | 12.0 | 26        | +/-  | 6.5       | 0.0              | NA  |

Sapir Zrihan, Clinical research assistant



Signature

19-Jul-22

Date

Frida Cohen, Clinical Manager

*Frida Cohen*

Signature

20-Jul-22

Date

Marcia Deutsch, CEO

*Marcia Deutsch*

Signature

21-Jul-22

Date