


QwikCheck™ Beads Batch Release Form

BATCH #	251121	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	December-22	LOT#	251121001	LOT#	251121002	LOT#	251121003		
For use with:	Conc.			Conc.			Conc	MSC	
SQA / QwikCheck™ gold Automated	47	+/-	6.6	26	+/-	5.2	0.0	0.0	
SQA-V Visualization	48	+/-	12.0	25	+/-	6.3	0.0	NA	
Hemocytometer	49	+/-	12.3	24	+/-	6.0	0.0	NA	
MAKLER	49	+/-	12.3	24	+/-	6.0	0.0	NA	
Fixed Coverslip	51	+/-	12.8	27	+/-	6.8	0.0	NA	

Sapir Zrihan, Clinical research assistant




Signature

6-Dec-21

Date

Frida Cohen, Clinical Manager



Signature

6-Dec-21

Date

Marcia Deutsch, CEO

MD

Signature

6-Dec-21

Date