

QwikCheck™ Low Range Beads Batch Release Form

BATCH #	270622	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	27-Mar-23	LOT#	270622001	LOT#	270622002	LOT#	270622003		
For use with:		Conc.			Conc.			Conc	MSC
SQA / QwikCheck™ gold Automated		16	+/-	6.4	3	+/-	2.0	0.0	0.0
SQA-V Visualization		16	+/-	6.4	4	+/-	2.6	0.0	NA
Fixed Coverslip		12	+/-	4.8	2	+/-	1.3	0.0	NA
Hemocytometer		18	+/-	7.2	NA	+/-	NA	0.0	NA

Victoria Goldenberg
Clinical research assistant

Victoria
Signature

29th June 2022
Date

Frida Cohen
Clinical Manager

Frida
Signature

29th June 2022
Date

Marcia Deutsch
CEO

Marcia
Signature

29th June 2022
Date