QwikCheck™ Liquefaction Kit Batch Release Form

QwikCheck™ Liquefaction Kit produced: DEC (Month) 2021 (Year)

Chymotrypsin activity = 1200 units/mg enzyme

Notes:
- **PASS criteria: Mean ≥1200units/mg enzyme**
- **Certificate of Analysis (Faizyme) for each Lot Number should be attached to this form**

Product insert	10	20	70
Vial labels	**1**000	400	1400
Box labels	10	20	70
QC labels	10	20	70
QC Approval date	1st June 21	1st June 21	1st June 21
Preparation date	2,3/Nov/21	13/Dec/21	23/Dec/21
Release date	7th Dec 2021	23rd Dec 2021	26th Dec 2021
Signature	**Victoria**	**Victoria**	**Victoria**

Batch number: ___210303_1221_____

Lot Number (Faizyme): ____**210303**_____

Exp. Date: ____**DEC**____ (**Month**) ____**2022**____ (**Year**)

Frida

_____ Clinical Research Assistant

__30th DEC 21__ Date